

INSPECTION OF SELECTED MEDICAL DEVICE ROUTINE CLEANING PROCEDURES: DETECTION OF RESIDUAL PROTEINS IN THE CLINICAL SETTING IN GAUTENG HOSPITALS

SJ JARDINE

A research dissertation submitted to the
Faculty of Health Sciences, University of the Witwatersrand, Johannesburg
fulfilment of the requirements for the degree
of
Master of Science in Nursing

Johannesburg, 2015

DECLARATION

I, SJ Jardine, declare that this research dissertation is my own work. It is being submitted for the degree of Master of Science (in Nursing) at the University of the Witwatersrand, Johannesburg. It has not been submitted before for any degree or examination at this or any other university.

Signature

.....day of 2015

Protocol Number: M130345

DEDICATION

I dedicate this work to all the nurses and nurse tutors who have upheld the nurses' pledge and have devoted their lives to improving patient care.

ACKNOWLEDGEMENTS

I would like to thank all my mentors over the years that have guided my life and nursing career (that includes you Malcolm).

Thank you also to the team of supervisors who have overseen this project who are all dedicated to improved patient care (Viv Herbert, Prof Duse and Shelly Schmollgruber).

A heartfelt thanks to SafMed (and Denise) for helping me develop and grow my skills and expand my career.

Thank you to the hospital groups private and public for allowing me to conduct this research, and of course to the staff who participated.

To my loved ones thank you for your patience.

To all of you I will always be grateful.

ABSTRACT

Background:

Patients undergo surgical and medical procedures daily in hospitals and clinics in South Africa to treat various health problems. Medical devices are used to perform these procedures. It is possible that dirty medical devices could transmit healthcare associated infections, so it is critical that medical devices are cleaned, disinfected or sterilized according to validated procedures. Medical devices should be visually inspected to verify that they are indeed clean. Not all patient soils are visible to the naked eye so cleaning should be verified using additional methods. One such method is to test the device for protein residues.

The aim of this study was to establish if selected hospitals in Gauteng (South Africa) have standard operating procedures for cleaning medical devices, if their cleaning procedures are based on international validated procedures and to investigate if selected medical devices have protein residuals left on them after undergoing routine cleaning procedures. The results of this study could be used to help establish South African medical device cleaning guidelines.

A descriptive, multiple case study design was utilised in order to understand the phenomenon of medical device cleaning within its real life context in 5 hospitals in Gauteng (South Africa). The researcher established if the hospitals had medical device cleaning standard operating procedures at a unit level. The researcher observed and documented how specific medical devices were routinely cleaned (namely; flexible gastroscope, Yankhauer suction nozzle, needle holder, diathermy forceps, laryngoscope blades and vaginal speculum). Those devices were then visually inspected for soil and photographed. A marked (traceable) Crile's forceps provided by the researcher was soiled with Browne's soil test (artificial test soil). The Crile forceps and the aforementioned

medical devices were then be swabbed with a sterile swab and that swab was tested for protein residues with a ninhydrin protein residual test.

Results

Only 3% percent of the staff at unit level was aware of medical device cleaning SOP's. Not all medical devices were cleaned in accordance with internationally validated cleaning procedures. The percentage of medical devices that tested positive for proteins varied per device and per hospital. Overall 16% percent of medical devices tested positive for residual proteins post routine cleaning. Only 12% of the medical devices that were kept moist tested positive for residual proteins, whereas 19% of the medical devices that were not kept moist tested positive for residual proteins. The commercially available Ninhydrin test kit (Browne Ltd, UK) was able to effectively detect residual proteins on medical devices in this research to a similar degree found in previous research. The findings in this research however suggest that the ninhydrin test method may not be ideal for testing gastroscopes for residual protein (a more sensitive test may be more effective), and that medical devices should be tested regularly for residual proteins as visually inspecting them for soils is simply unreliable.

Conclusion

Medical devices should be decontaminated according to hospital SOP's that are based on validated cleaning procedures and MIFU in order for them to be cleaned effectively. Cleaning of medical devices should be verified using residual protein tests as visual inspection alone is not effective enough. South African guideline should be established for cleaning of medical devices that are in line with internationally validated procedures, MIFU and that advocate the use of protein residual tests to verify cleaning efficacy.

TABLE OF CONTENTS

DECLARATION	ii
DEDICATION	iii
ACKNOWLEDGEMENTS	iv
ABSTRACT	v
TABLE OF CONTENTS	vii
LIST OF FIGURES	xii
LIST OF TABLES	xv
LIST OF ACRONYMS	xv

CHAPTER ONE: OVERVIEW OF THE STUDY		
1.1	INTRODUCTION	1
1.2	BACKGROUND REGARDING MEDICAL DEVICE DECONTAMINATION	1
1.2.1	Manual and automated cleaning methods	5
1.2.2	Standard operating procedures for cleaning	6
1.2.3	Education and training	6
1.2.4	Medical device design	7
1.3	PROBLEM STATEMENT	9
1.4	RESEARCH QUESTIONS	9
1.5	PURPOSE OF THE STUDY	9
1.6	RESEARCH OBJECTIVES	10
1.7	SIGNIFICANCE OF THE STUDY	10
1.8	RESEARCHER'S ASSUMPTIONS	12
1.8.1	Meta-theoretical Assumptions	12
1.8.2	Theoretical Assumptions	14
1.8.3	Methodological Assumptions	17
1.9	OVERVIEW OF METHODOLOGY	18
1.9.1	Research Design	18
1.9.2	Research Methods	18
1.9.3	Pilot Study Phase 1	19
1.9.4	Population	20
1.9.5	Sample and sampling	20
1.9.6	Data Collection	21
1.9.7	Structured observation check list (Instrument)	21

1.9.8	Data Analysis	21
1.10	VALIDITY AND RELIABILITY OF THE STUDY	22
1.11	ETHICAL CONSIDERATIONS	22
1.12	SUMMARY	23

CHAPTER TWO: LITERATURE REVIEW		
2.1	INTRODUCTION	24
2.2	BACKGROUND	24
2.3	TRANSMISSION OF DISEASE VIA CONTAMINATED MEDICAL DEVICES	25
2.4	THE BURDEN OF HEALTHCARE ASSOCIATED INFECTIONS	27
2.5	WHY IS MEDICAL DEVICE CLEANING SO IMPORTANT	28
2.6	FACTORS THAT AFFECT MEDICAL DEVICE CLEANING	29
2.6.1	Manufacturer's instructions for use (MIFU)	29
2.6.2	Policies and Procedures	31
2.6.3	Education and Training of Staff	32
2.6.4	Keeping instruments moist	33
2.6.5	Design of Instruments	34
2.6.6	Manual and Automated Cleaning	39
2.7	HOW TO VERIFY CLEANING WAS DONE EFFECTIVELY	40
2.8	VISUAL INSPECTION	40
2.9	PROTEIN RESIDUAL TESTS	42
2.9.1	Ninhydrin Method	42
2.9.2	Modified ortho-phthalic dialdehyde (OPA) method	43
2.9.3	Biuret Method	43
2.10	SUMMARY	43

CHAPTER THREE: RESEARCH METHODOLOGY		
3.1	INTRODUCTION	44
3.2	RESEARCH METHODOLOGY	44
3.3	RESEARCH DESIGN	44
3.3.1	Descriptive Research	45
3.3.2	Case Study	45
3.4	PILOT STUDY	48
3.4.1	Research setting	49

3.4.2	Population	49
3.4.3	Sample and sampling	49
3.4.4	Data collection	50
3.4.5	Structured Observation Checklist (SOCL) tool for data collection	50
3.4.6	Data collection procedure	51
3.5	RESEARCH SETTING	57
3.6	POPULATION	58
3.7	SAMPLE AND SAMPLING	59
3.7.1	Sample size	60
3.8	DATA COLLECTION	61
3.8.1	Structured Observation Checklist (SOCL) tool for data collection	61
3.8.2	Validity and reliability of the structured observation checklist (Instrument)	63
3.8.3	Data Collection Procedure	64
3.	VALIDITY AND RELIABILITY OF STUDY	70
3.9.1	Construct Validity	70
3.9.2	Internal Validity	71
3.9.3	External validity	72
3.9.4	Reliability	72
3.10	ETHICAL CONSIDERATIONS	72
3.11	SUMMARY	74

CHAPTER FOUR: DATA ANALYSIS AND RESULTS		
4.1	INTRODUCTION	75
4.2	APPROACH TO DATA ANALYSIS	75
4.3	RESULTS OF THE PILOT STUDY	76
4.3.1	Research setting	76
4.3.2	Number of protein residual tests done at 1 hospital	76
4.3.3	Cleaning observed or described	77
4.3.4	Where devices were cleaned	78
4.3.5	Designation of the person cleaning	79
4.3.6	Cleaning SOP	80
4.3.7	Cleaning Method: Manual or Automated	80
4.3.8	Type of Detergent	81
4.3.9	Dilution of Detergent and Water	82

4.3.10	Kept Moist	82
4.3.11	Friction	83
4.3.12	Cleaning accessories	84
4.3.13	Flush lumen manual cleaning	85
4.3.14	Macroscopically Visually Clean	85
4.3.15	Gastrosopes	86
4.3.16	Cleaned gastrosopes tested positive for residual proteins	87
4.3.17	Medical devices positive for residual proteins overall	88
4.3.18	Positive for proteins, per device	88
4.3.19	Positive protein test reaction time	89
4.3.20	Pattern matching: Positive for proteins: washed manually	90
4.3.21	Pattern matching: Positive for proteins: kept moist	91
4.3.22	Pattern Matching: Positive for proteins: Yankhauer suction lumen flushed vs. not flushed	92
4.3.23	Pattern Matching: Visually Clean: Positive for proteins	93
4.3.24	Pattern Matching: Visually Dirty: Positive for proteins	94
4.4	RESULTS OF MAIN STUDY	95
4.4.1	Research Setting	95
4.4.2	Number of Protein Residual Tests done per hospital	95
4.4.3	Number of Protein Residual Tests done per hospital; per device	96
4.4.4	Cleaning observed or described	97
4.4.5	Where devices were cleaned	98
4.4.6	Designation of the person cleaning	99
4.4.7	Cleaning SOP	100
4.4.8	Cleaning Method: Manual or Automated	101
4.4.9	Type of Detergent	101
4.4.10	Dilution of Detergent and Water	102
4.4.11	Kept Moist	103
4.4.12	Automated Cleaning: Pre clean	104
4.4.13	Washer Manufacturer	104
4.4.14	Correct Tray	105
4.4.15	Shadowing	106
4.4.16	Overloaded	107
4.4.17	Friction	107
4.4.18	Cleaning accessories	108

4.4.19	Flush lumen manual cleaning	109
4.4.20	Flush lumen automated cleaning	110
4.4.21	Macroscopically Visually Clean	110
4.4.22	Cleaning of Gastrosopes: Leak Test Available	111
4.4.23	Cleaning of Gastrosopes: Leak Test Performed	112
4.4.24	Cleaning of gastrosopes: cleaning steps performed when cleaning manually	113
4.4.25	Cleaning of gastrosopes: Manual Clean Time	114
4.4.26	Cleaning of gastrosopes: Manual vs. Automated	114
4.4.27	Dirty gastrosopes tested positive for residual proteins	115
4.4.28	Cleaned gastrosopes tested positive for residual proteins	116
4.4.29	Pattern matching gastrosopes: positive for proteins: washed manually	116
4.4.30	Pattern matching gastrosopes: positive for proteins: washed in automated washer	117
4.4.31	Pattern matching gastrosopes: clean-time positive for proteins	118
4.4.32	Medical devices positive for residual proteins overall	118
4.4.33	Overall average positive for proteins per hospital	119
4.4.34	Positive for proteins, per device	120
4.4.35	Positive Protein Test Reaction Time	120
4.4.36	Pattern Matching: Washed manually: positive for residual proteins	121
4.4.37	Pattern Matching: Washed in an automated washer: positive for residual proteins	122
4.4.38	Pattern Matching: Positive for proteins: kept moist	122
4.4.39	Pattern Matching: Positive for proteins: Yankhauer suction lumen flushed vs. not flushed	123
4.4.40	Pattern Matching: Visually Clean: Positive for proteins	124
4.4.41	Pattern Matching: Visually Dirty: Positive for proteins	125
4.5	SUMMARY	125

CHAPTER FIVE: SUMMARY, DISCUSSION OF RESULTS, LIMITATIONS, CONCLUSIONS AND RECOMMENDATIONS		
5.1	INTRODUCTION	126
5.2	PURPOSE OF THE STUDY	126
5.3	RESEARCH OBJECTIVES	126
5.4	METHODOLOGY	127
5.5	SUMMARY OF THE MAIN FINDINGS	128

5.6	LIMITATIONS OF THE STUDY	139
5.7	RECOMMENDATIONS	140
5.7.1	Recommendations for healthcare education	140
5.7.2	Recommendations for clinical practice	141
5.7.3	Recommendations for further research	142
5.8	CONCLUSION	142

LIST OF REFERENCES	144
---------------------------	-----

APPENDICES		
APPENDIX A	CASE STUDY PROTOCOL	153
APPENDIX B	INFORMATION LETTER FOR PARTICIPANTS	165
APPENDIX C	INFORMED CONSENT FORM	167
APPENDIX D	STRUCTURED OBSERVATION CHECK LIST PILOT STUDY	168
APPENDIX E	STRUCTURED OBSERVATION CHECK LISTS MAIN STUDY GASTROSCOPES	170
APPENDIX F	STRUCTURED OBSERVATION CHECK LISTS MAIN STUDY MEDICAL DEVICES	171
APPENDIX G	APPROVAL FROM GAUTENG PROVINCIAL HOSPITAL 1	172
APPENDIX H	APPROVAL FROM GAUTENG PROVINCIAL HOSPITAL 2	173
APPENDIX I	APPROVAL FROM HUMAN RESEARCH ETHICS COMMITTEE	174
APPENDIX J	APPROVAL TO CONDUCT RESEARCH PRIVATE HEALTHCARE GROUP 1	176
APPENDIX K	APPROVAL TO CONDUCT RESEARCH PRIVATE HEALTHCARE GROUP 2	177
APPENDIX L	APPROVAL TO CONDUCT RESEARCH PRIVATE HEALTHCARE GROUP 3	178
APPENDIX M	APPROVAL POSTGRADUATE COMMITTEE	180

LIST OF FIGURES

Figure		Page
1.1	Decontamination lifecycle model	4
3.1	Purple colour change	53
3.2	Gastroscope biopsy port	53
3.3	Brush tip	53

3.4	Tip of brush in vial	53
3.5	Vial in incubator	53
3.6	Yankhauer suction	54
3.7	DispoClean brush	54
3.8	Tip of brush in vial	54
3.9	Needle holder	55
3.10	Diathermy forceps	55
3.11	Laryngoscope blade	56
3.12	Vaginal speculum	56
3.13	Crile's forceps	57
3.14	The data collection procedure flow chart: pilot study and main study	64
3.15	Gastroscope biopsy port	66
3.16	Brush tip	66
3.17	Tip of brush in vial	66
3.18	Vial in incubator	66
3.19	Portion of yankhauer suction swab tested for residual protein	67
3.20	Portion of needle holder swab tested for residual protein	67
3.21	Portion of diathermy forceps swab tested for residual protein	68
3.22	Portion of laryngoscope swab tested for residual protein	69
3.23	Portion of vaginal speculum swab tested for residual protein	69
3.24	Portion of Crile's forceps swab tested for residual protein	70
4.1	Research setting	76
4.2	Percentage of medical device cleaning observed vs. described	78
4.3	Where in the hospital medical devices were cleaned pilot study	79
4.4	Designation of the person cleaning	79
4.5	Percentage not aware of medical device cleaning SOP's	80
4.6	Percentage of devices cleaned manually vs. automated cleaning pilot	81
4,7	Type of detergent used to clean medical devices pilot study	82
4.8	Detergent diluted according to MIFU pilot study	82
4.9	Medical devices kept moist until cleaned	83
4.10	Medical devices cleaned using friction	84
4.11	Cleaning accessories	84
4.12	Percentage of lumens of yankhauer suction flushed	85
4.13	Percentage of devices visually clean	86
4.14	Cleaned gastroscopes tested positive for protein	87

4.15	Overall percentage of devices positive for proteins pilot study	88
4.16	Positive protein test reaction time pilot study	90
4.17	Washed manually positive for proteins pilot study	91
4.18	Percentage of devices kept moist positive for proteins pilot study	92
4.19	Percentage of yankhauer suctions positive for proteins: lumen flushed vs. lumens not flushed	93
4.20	Visually clean but positive for proteins	94
4.21	Visually dirty but positive for proteins	94
4.22	Research setting main study	95
4.23	Number of residual protein tests done per hospital main study	96
4.24	Percentage of medical device cleaning observed vs described main study	98
4.25	Area in the hospital where medical devices were cleaned main study	99
4.26	Designation of the staff cleaning medical devices main study	100
4.27	Percentage aware of a cleaning SOP main study	100
4.28	Percentage of devices cleaned manually vs. automated cleaning main study	101
4.29	Type of detergent used to clean medical devices main study	102
4.30	Detergent diluted according to MIFU main study	103
4.31	Percentage of medical devices kept moist until cleaned main study	103
4.32	Percentage of devices pre cleaned prior to automated cleaning	104
4.33	Automated washer manufacturers	105
4.34	Percentage of time correct trays was used in automated washer	105
4.35	Percentage of medical device shadowed in the automated washer	106
4.36	Percentage of devices washed in overloaded automated washer	107
4.37	Percentage medical devices cleaned using friction main study	108
4.38	Cleaning accessories used main study	109
4.39	Percentage of lumens of yankhauer suction flushed during manual cleaning main study	109
4.40	Percentage of lumens of yankhauer suction flushed during automated cleaning main study	110
4.41	Percentage of devices visually clean main study	111
4.42	Percentage of times leakage test available	112
4.43	Percentage of time leakage test performed	112
4.44	Percentage of times scope cleaning steps performed for manual cleaning	113

4.45	Time taken to clean gastroscope manually	114
4.46	Number of gastroscopes cleaned manually and in an automated washer	115
4.47	Number of dirty gastroscopes tested positive for proteins	115
4.48	Number of cleaned gastroscopes tested positive for proteins	116
4.49	Gastroscopes washed manually positive for proteins	117
4.50	Gastroscopes washed in an automated washer positive for proteins	117
4.51	Number of gastroscopes positive for proteins when cleaned for less than one minute and greater than one minute	118
4.52	Overall percentage of device positive for proteins main study	119
4.53	Overall percentage of device positive for proteins, per hospital main study	119
4.54	Positive protein test reaction time main study	121
4.55	Washed manually positive for proteins main study	121
4.56	Washed in automated washer: positive for proteins	122
4.57	Percentage of devices kept moist positive for proteins main study	123
4.58	Percentage of yankhauer suction flushed and not flushed positive for proteins	124
4.59	Medical devices that were visually clean but tested positive for proteins	124
4.60	Medical devices that were visually dirty but tested positive for proteins	125
5.1	Stained swab	140

LIST OF TABLES

Table		Page
3.1	Sample size pilot study	50
3.2	Sample size main study	60
4.1	Number of residual protein tests done per medical device	77
4.2	Number of medical devices tested positive per device pilot study	89
4.3	Number of residual protein tests intended vs. actual tests done per device per hospital	97
4.4	Number of medical devices tested and number positive for proteins per device	120

LIST OF ACRONYMS

AAMI	Association for the Advancement of Medical Instrumentation
AORN	Association of Operating Room Nurses
CDC	Centres for Disease Control and Prevention
CJD	Creutzfeldt-Jakob Disease
CSSD	Central Sterile Services Departments
CSSD TECH	Central Sterile Services Departments Technician
EN	Enrolled Nurse
ENA	Enrolled Nursing Assistant
FDA	Food and Drug Administration
HCAI	Healthcare Associated Infection
HIV	Human Immunodeficiency Virus
ICU	Intensive Care Unit
ISO	International Standards Organisation
MIFU	Manufacturer's Instruction for Use
NHS	National Health System
NON HCW	Non Health Care Worker
OPA	Ortho-Phthalic Dialdehyde
PCD	Process Challenge Device
RN	Registered Nurse
SANAC	South African National Aids Council
SOCL	Structured Observation Check List
SOP	Standard Operating Procedure
STI	Sexually Transmitted Infection
STN	Student Nurse
TB	Tuberculosis
TSE	Transmissible Spongiform Encephalopathies
vCJD	variant Creutzfeldt-Jakob Disease
WHO	World Health Organisation